

RESEARCH ARTICLE

Open Access



Effectiveness of photofunctionalized titanium alloy on osseointegration in rats with type 2 diabetes

Shengdao Jin, Yuji Yamamoto^{*}, Yoshifumi Harada, Sho Kaneko, Kazuki Oishi and Yasuyuki Ishibashi

Abstract

Background: Ultraviolet (UV) light-mediated photofunctionalization improves the osseointegration of pure titanium and titanium alloy (Ti6Al4V). However, little is known about the effect of UV irradiation on Ti6Al4V, used frequently in orthopedic surgery, in diabetic patients. We examined the effect of UV irradiation on Ti6Al4V in rats with type 2 diabetes.

Methods: Cylinder Ti6Al4V implants were used. Half the animals were Sprague Dawley rats (the control group), and the other half were Spontaneously Diabetic Torii fatty rats (the diabetes mellitus model). For radiological analysis, bone density was observed and calculated using 3D microcomputed tomography. Histological analysis was performed to calculate the bone–implant contact (BIC) ratio. We used Pearson correlation to analyze the correlation between average blood glucose level and BIC ratio, and between average blood glucose level and bone volume (BV) ratio.

Results: In the UV light-treated group, the BIC ratios of the normal and diabetic rats increased significantly compared with those in the untreated group at 2 weeks; at 4 weeks, the BIC ratio of the diabetic rats increased significantly, but there was no significant increase in the control animals. In both the control and diabetic groups, there was no significant difference in the BV ratios between the UV-treated and untreated implants at 2 or 4 weeks. The average blood glucose level in the 4-week group negatively correlated with the BIC and BV ratios. The average blood glucose level in the UV-treated group negatively correlated with the BIC ratio.

Conclusion: Photofunctionalization of Ti6Al4V implants may promote osseointegration in the early stages in rats with type 2 diabetes.

Keywords: Ultraviolet, Ti6Al4V, Diabetes mellitus, Osseointegration, Photofunctionalization

Background

As the world's population ages, the number of total hip arthroplasties (THA) and total knee arthroplasties (TKA) are increasing rapidly worldwide [1–3]. Titanium alloy (Ti6Al4V) implants are often used as prosthetic devices and one of the main materials used for fracture fixation in orthopedic surgery because of their advantages,

which include corrosion resistance, good biocompatibility, high stiffness, and a high weight-bearing tolerance [4–6]. Although there is a growing demand for such implant surgeries in an aging society, the number of patients with poor bone quality has increased, resulting in poor bone–implant integration, which often causes serious complications. Improving the speed and strength of bone–titanium integration remains a long-term challenge in the field of orthopedics [7, 8].

Diabetes mellitus (DM) is a common comorbidity associated with adverse outcomes in surgical patients. In 2019, the International Diabetes Federation announced

*Correspondence: yuji1112@hirosaki-u.ac.jp

Department of Orthopaedic Surgery, Hirosaki University Graduate School of Medicine, 5 Zaifu-cho, Hirosaki, Aomori 036-8562, Japan



© The Author(s) 2022. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

that there were approximately 463 million diabetic patients worldwide, and this number is expected to reach 700.2 million by 2045 [9]. DM is a risk factor for poor osseointegration, surgical site infections, aseptic loosening, dislocation, periprosthetic fracture, readmission, and mortality after patients have undergone either THA or TKA [10–12]. Clinical studies demonstrate that patients with type 2 diabetes have higher implant failure rates than nondiabetic patients [13, 14], and diabetic animal models demonstrate significant reductions in osseointegration parameters, particularly in the percentage of bone–implant contact (BIC) [15–17].

Recently, it has been shown that irradiating implants with ultraviolet (UV) light improve associated cell growth and osseointegration [18, 19]. This innovative technology enhances osseointegration, and the BIC ratio of photofunctionalized titanium implants increased to a near-maximum level of 98.2% in a rat model, 1.9 times that of the untreated implants at 4 weeks. The effect of photofunctionalization on Ti6Al4V surfaces was demonstrated *in vitro* to enhance both bioactivity and osteoconductivity [20]. After UV irradiation, the carbon content of Ti6Al4V surfaces decreases significantly, and the hydrophilic surface becomes more hydrophilic (contact angles decreased from 72.3 to 6.0 degrees) [21]. Further, photofunctionalization is effective in improving the survival rate and stability of Ti6Al4V screws under loading conditions [22].

Photofunctionalization accelerated and enhanced levels of osseointegration, and overcame impaired osseointegration, on pure titanium in a rat model of type 2

diabetes [23]. However, the effect of photofunctionalized Ti6Al4V on *in vivo* bone histomorphometric parameters, such as the BIC ratio in diabetic patients, is still unclear. We aimed to examine the effect of photofunctionalized Ti6Al4V on osseointegration in rats with type 2 diabetes.

Methods

Animal experiments were performed to demonstrate the effects of photofunctionalized Ti6Al4V on osseointegration in rats with type 2 diabetes. In this *in vivo* study, Ti6Al4V implants were inserted into rat femurs, the rats were killed after 2 or 4 weeks and radiological analysis was performed using microcomputed tomography, followed by histological analysis using undecalcified specimens. The study protocol (ethical code number: M19007) was approved by the Animal Research Committee of Hirosaki University, and all experiments were performed according to the Rules for Animal Experimentation of University.

Photofunctionalization of Ti6Al4V implants

The cylinder implants were made from Ti6Al4V (diameter 2 mm, length 8 mm) and provided by B. Braun Aesculap Japan Co., Ltd (Tokyo, Japan). Half of the implants were treated with UV irradiation for 15 min using a photodevice (TheraBeam® affinity; Ushio Inc., Tokyo, Japan) at an intensity of 3 mW/cm² (Fig. 1A–C). The light source mounted in the TheraBeam affinity was a low-pressure mercury (Hg) lamp, which emitted predominantly UV light of 185-nm and 254-nm. The implants were divided into two groups: Ti6Al4V without UV irradiation




Fig. 1 Photofunctionalization and implant. **a** Cylinder Ti6Al4V implants (diameter, 2 mm; length, 8 mm). **b** Implants were subjected to ultraviolet irradiation for 15 min

(untreated group) and Ti6Al4V with UV irradiation (UV-treated group).

Animals

The Spontaneously Diabetic Torii (SDT) fatty rats (average adult weight of 360.7 ± 18.4 g) reach a high blood glucose level (approximately 200 mg/dL) at 6 weeks of age, and a very high level (approximately 400 mg/dL) at 8 weeks, and thus are a very mature model for type 2 diabetes [24–26].

Sprague Dawley (SD) rats (average adult weight of 266.7 ± 13.5 g) were used for the control group and SDT fatty rats for the DM model. All rats were 8-week-old males and purchased from the same company (CLEA Japan, Inc., Tokyo, Japan). The rats were maintained in bracket cages and fed a standard laboratory diet and were able to access water *ad libitum* under temperature-, humidity-, and lighting-controlled conditions.

A total of 20 rats were divided into four groups, with five in each group as follows: Group I: SD rats implanted for 2 weeks; Group II: DM rats implanted for 2 weeks; Group III: SD rats implanted for 4 weeks; and Group IV: DM rats implanted for 4 weeks.

Surgery


The rats were anesthetized with 1–2% isoflurane. Both hind limbs were shaved, and the incision area (from the distal femur to the knee) was wiped with alcohol before the skin and fascia were opened separately. The flat aspect of each distal femur was exposed and used for implantation. The bilateral distal femurs were drilled using a 2-mm diameter drill. UV-treated implants were inserted into the right femur holes, and untreated implants were inserted into the left femur holes (Fig. 2). After implant placement, the skin and fascia were closed with stitches. At either 2 or 4 weeks after surgery, the rats were euthanized by drawing more than 5 mL of blood directly from the heart, and the femurs were harvested.

Blood glucose analysis

To confirm the establishment of normal and diabetic rats, blood glucose levels were measured just before implantation (0 weeks) and every 2 weeks after the operation until kill. Blood was collected from the tail of the rats, and blood glucose levels were measured using a blood glucometer (Experimental Animal Glucometer SUGL-001, ForaCare Japan, Japan).

Radiological analysis

The specimens were analyzed using microcomputed tomography (Scan Xmate-L090, Comscantecno Co., Ltd., Japan). The imaging conditions were as follows: voltage, 80 kV; current, 100 μ A; magnification, 4.942 times;



resolution, 20.234 μ m/pixel; and slice thickness, 20.234 μ m. Three-dimensional bone morphometric analysis was performed using the TRI-3D-BON software (TRI/3D-BON, RATOC System Engineering Co., Ltd., Japan). The bone volume (BV) ratio was defined as the ratio of the mineralized BV within 100 μ m from the implant surface. The BV ratio was calculated as the bone occupancy in the area of interest divided by the total area of interest, multiplied by 100.

Histological analysis

The specimens were fixed in 10% buffered formalin and analyzed using microcomputed tomography (Scan Xmate-L090, Comscantecno Co., Ltd., Japan). Specimens were embedded in methyl methacrylate without decalcification [27]. The implants were left in situ for histological analysis. Embedded specimens were cut along the long axis of the implants using a microtome (EXAKT, Norderstedt, Germany).

Each 30–40 μ m section was stained green with Villanueva–Goldner and examined by light microscopy (BZ-X700; Keyence Corp., Japan) to evaluate the bone area. For each histological slice, the BIC ratio for each group was calculated using digital image analysis software (ImageJ version 1.48). The BIC ratio was calculated as the length of the bone in direct contact with the surface of the implant divided by the total length of the implant, multiplied by 100 (Fig. 2). The bone in direct contact was defined as the interface at which the bone tissue was

located within 20 μm of the implant surface without the intervention of soft tissue.

Statistical analyses

The Mann–Whitney U test was performed to determine differences in blood glucose levels. The Wilcoxon signed-rank test was performed to determine differences in the BIC and BV ratios between the UV-treated and untreated groups at 2 or 4 weeks, respectively. The Mann–Whitney U test was also performed to determine differences in BIC and BV ratios between 2 and 4 weeks in each group. Statistical analyses were performed using SPSS (v 21.0; IBM), and *p*-values of <0.05, were considered significant.

The Pearson correlation analysis method was used to analyze the correlation between mean blood glucose level and BIC ratio and between mean blood glucose level and BV ratio. Correlation analysis of all the data was performed, and the data were analyzed after different groupings. Correlation analyses were performed using SPSS (v 21.0; IBM), and *p*-values of <0.05 were considered to indicate correlation. When *r* is >0, the two variables are positively correlated, and when *r* <0, the two variables are negatively correlated. When $|r| \geq 0.8$, the two variables were highly correlated; when $0.5 \leq |r| < 0.8$, the correlation was moderate; when $0.3 \leq |r| < 0.5$, there was a low correlation; and when $|r| < 0.3$, the correlation between the two variables was very weak and was regarded as uncorrelated.

Results


Blood glucose analysis

Based on the blood glucose analysis, the blood glucose of Group I (SD rats implanted for 2 weeks) was determined to be 115.3 ± 15.0 mg/dL just before surgery (0 weeks) and 110.3 ± 14.9 mg/dL at 2 weeks after surgery (2 weeks) (Fig. 3a). The blood glucose of Group II (DM rats implanted for 2 weeks) was 384.4 ± 62.6 mg/dL (0 weeks) and 533.6 ± 71.1 mg/dL (2 weeks) (Fig. 3a). In the 4-week groups, the blood glucose of Group III (SD rats implanted for four weeks) was 115.2 ± 13.5 mg/dL (0 weeks), 110.2 ± 11.0 mg/dL (2 weeks), and 136.6 ± 10.6 mg/dL (4 weeks) after surgery (Fig. 3b). The blood glucose of Group IV (DM rats implanted for 4 weeks) was 364.4 ± 55.0 mg/dL (0 weeks), 489 ± 45.4 mg/dL (2 weeks), and 521.2 ± 49.0 mg/dL (4 weeks) (Fig. 3b).

At different time points, the blood glucose level of diabetic rats in each group was significantly higher (greater than 300 mg/dL) than that found in the healthy rats; thus, the diabetic model was established (Fig. 3a, b).

Radiological analysis

In the SD groups, the mean BV ratio was $60.8\% \pm 9.3\%$ for untreated rats and $59.2\% \pm 5.3\%$ for UV-treated rats




at 2 weeks; and $67.2\% \pm 5.6\%$ for untreated rats and $61.7\% \pm 10.0\%$ for UV-treated rats at 4 weeks (Figs. 4 and 5). There were no significant differences between the untreated and UV-treated groups at 2 or 4 weeks (*p* = 0.715 and *p* = 0.465, respectively).

In the DM group, the mean BV ratio was $54.4\% \pm 13.1\%$ for untreated rats and $51.7\% \pm 9.1\%$ for UV-treated rats at 2 weeks; and $49.2\% \pm 6.0\%$ for untreated rats and $46.1\% \pm 6.8\%$ for UV-treated rats at 4 weeks. Moreover, there were no significant differences between the untreated and UV-treated groups at 2 or 4 weeks (*p* = 0.686 and *p* = 0.068, respectively).

Histological analysis

In the SD groups, the mean BIC ratio was $47.9\% \pm 8.2\%$ for untreated rats and $67.0\% \pm 5.1\%$ for UV-treated rats at 2 weeks; and $75.5\% \pm 12.9\%$ for untreated rats and




79.9% ± 4.8% for UV-treated rats at 4 weeks. The BIC ratio increased significantly at 4 weeks compared with that at 2 weeks in both untreated and UV-treated groups ($p=0.016$ and $p=0.008$, respectively). There was a significant difference in the BIC ratio between the untreated and UV-treated groups at 2 weeks ($p=0.043$); however, there was no significant difference at 4 weeks ($p=0.345$).

In the DM group, the mean BIC ratio was 40.4% ± 8.3% for untreated rats and 58.1% ± 5.7% for UV-treated rats at 2 weeks and 64.2% ± 4.2% for untreated rats and 70.2% ± 8.0% for UV-treated rats at 4 weeks. The BIC ratio increased significantly at 4 weeks compared with that at 2 weeks in the untreated groups ($p=0.008$); however, there was no difference in the UV-treated groups ($p=0.056$). There were significant differences in the BIC ratio between the untreated and UV-treated groups at both 2 and 4 weeks ($p=0.043$ and $p=0.043$, respectively) (Figs. 6 and 7).

Correlation analysis

Data analysis in all groups (Groups I, II, III, and IV) showed a slight negative correlation between mean blood glucose level and the BIC ratio ($r = -0.368$, $p = 0.023$); however, there was no correlation between mean blood glucose level and BV ratio ($r = -0.220$, $p = 0.184$). Data were divided into the 2-week (Groups I and II) and 4-week (Groups III and IV) groups for correlation analysis. The average blood glucose level in the 4-week group was moderately negatively correlated with the BIC ratio



($r = -0.574$, $p = 0.008$), and there was also a slight negative correlation between mean blood glucose level and BV ratio ($r = -0.477$, $p = 0.034$). Similarly, when data were divided into the UV-treated and untreated groups for analysis, the average blood glucose level in the UV-treated group was moderately negatively correlated with the BIC ratio ($r = -0.586$, $p = 0.008$); however, there was no correlation between mean blood glucose level and BV ratio ($r = -0.302$, $p = 0.209$).

Discussion

We demonstrated that photofunctionalization can promote early phase osseointegration of Ti6Al4V in both type 2 diabetic rats and normal rats. The BIC ratios of surfaces at 2 and 4 weeks were significantly higher in the UV-treated group than in the untreated groups in diabetic rats. However, in normal rats, there was a significant difference between the UV-treated and untreated groups only at 2 weeks; there was no significant difference at 4 weeks.





Fig. 6 Light microscope images (magnification $\times 4$ and $\times 10$, higher-magnification images of the boxed areas). Light microscope images at 2 and 4 weeks after the implantation of the diabetes mellitus (DM) group. The micrographs show the bone response for the untreated (a, c, e, g) and ultraviolet (UV)-treated implants (b, d, f, h). A sample of the 2-week untreated group is shown in a and c, and sample from the 2-week UV-treated group is shown in b and d. One sample of the 4-week untreated group is shown in e and g, and a sample of the 4-week UV-treated group is shown in f and h. Scale bar: 100 μm



Currently, limited research has been conducted on the photofunctionalization of Ti6Al4V implants in diabetic models. A previous study demonstrated that promoting osseointegration promoted bone integration of pure titanium in diabetic rats [23]. It reported that the integration of the UV-irradiated implants at 2 weeks was 80% higher than that of the control group [23]. Another clinical study showed that the healing time for implant stability in the moderately and poorly controlled diabetic groups was approximately twice as long as that in nondiabetic and well-controlled diabetes groups [28].

Recently, the biological aging phenomenon of titanium has been demonstrated to be a limitation. A reduced BIC ratio is reportedly caused by biological aging secondary to time-dependent biological degradation of the Ti surface [6]. Ti surfaces constantly absorb hydrocarbons from the atmosphere, together with water and cleaning solutions after the implants are manufactured [29]. The amount of carbon on the Ti surface affects the initial affinity of osteoblasts and amount of bone–Ti integration [6, 30, 31]. Additionally, the absorption of hydrocarbons leads to an

increase in hydrophobia on the implant surface. Surface wettability is an important property for cell behavior, and by definition, cell attachment onto hydrophobic surfaces tends to be weaker than that onto hydrophilic surfaces [32–34]. Our previous study using the same implants as this study demonstrated that the amount of carbon and the contact angle on implants were significantly reduced after UV irradiation [21]. These changes of implant surface properties after UV irradiation enhance establishment of osseointegration in the early healing stage.

After UV irradiation, the osteoblasts on the surface of Ti6Al4V significantly increased by 80%–100% compared with those in the untreated group, thus enhancing the bioactivity and bone conductivity of Ti6Al4V [20]. UV irradiation can also enhance the osteogenesis around the implant, increase bone deposition on the titanium surface, and improve the sealing and support of the marginal bone [35]. Compared with the untreated implant, photofunctionalization of Ti6Al4V induced denser cortical bone formation and more rigid bone connection [23]. Similarly, this study demonstrated that photofunctionalization can promote early phase osseointegration of Ti6Al4V in both type 2 diabetic rats and normal rats. The fact that there was a significant difference in BIC between the UV-treated and the untreated groups and no significant difference in BV ratio was consistent with previous studies. While BIC evaluated new bone on the surface of the implant, BV ratio evaluated the proportion of bone tissue within 100 μ m around the implant; it suggests that the photofunctionalization effect is strong near the surface of the implant.

Early osseointegration is essential because a low BIC rate in the early phase can very easily lead to implant failure. Therefore, we evaluated the effect of photofunctionalization on Ti6Al4V surfaces in the early healing stage (up to 4 weeks). Increased hydrophilicity, which is one of the effects of UV irradiation, is reportedly observed at the implant surface after irradiation, but returns to its original state after 4 weeks [21]. Therefore, the effects of UV irradiation may not persist after 4 weeks. However, in this study, the BIC ratio of the diabetic rats in the UV-treated groups was approximately 10% lower than that in the control rats at 4 weeks. Further study is needed to definitively determine the lasting effects of photofunctionalization on Ti6Al4V in diabetics.

The correlations between mean blood glucose and BIC ratio and between mean blood glucose and BV ratio were also analyzed. There was no correlation between average blood glucose, BIC ratio, and BV ratio when data were analyzed, according to the 2-week and 4-week groups. However, the results of the 4-week group revealed a negative correlation between blood glucose and bone integration rate, and blood glucose and BV rate. If data

were divided into UV-treated and untreated groups, the UV-treated group showed a negative correlation between blood glucose and BIC rate. Therefore, we believe that for a specific group, blood glucose will affect the BIC ratio or BV ratio, but further research is needed to confirm this.

This study had several limitations. First, there is no implant biomechanical test to evaluate the biomechanical strength of bone–implant integration, although studies have shown that the push-in values of photo-functionalized Ti and Ti6Al4V implants are significantly higher than those of control implants [18, 23]. Second, each group contained only five samples, making it difficult to draw statistical conclusions. Post hoc power analysis indicated that five specimens provided a power of 0.70 to detect the difference of BIC ratio between UV-treated and untreated groups (effect size = 1.5, $\alpha = 0.05$). Third, to analyze bone regeneration around the implant, no fluorescence staining was performed. This method can be used as a dynamic evaluation method for bone regeneration.

Conclusions

UV irradiation can promote the early osseointegration of Ti6Al4V in diabetic models. Blood glucose level, UV radiation, and the time after implantation are all important factors that affect osseointegration. The strict control of blood glucose levels in patients with diabetes will help bone–implant integration. UV irradiation can be applied to orthopedic implants for diabetic patients, to promote early load-bearing and potentially reduce the implant failure rate.

Abbreviations

UV: Ultraviolet; DM: Diabetes mellitus; Ti: Titanium; Ti6Al4V: Titanium alloy; BIC: Bone–implant contact; BV: Bone volume; TV: Tissue volume.

Acknowledgements

We would like to thank Editage (www.editage.com) for English language editing.

Author's contributions

SJ contributed to acquisition of data, analysis and interpretation of data, and drafting the article. YY contributed to study conception and design, data, analysis and interpretation of data, and drafting and revising the article. YH, SK, and KO analyzed and interpreted data and drafted and revised the article. YI contributed to study conception and design, analysis and interpretation of data, and drafting and revising the article. All authors read and approved the final manuscript.

Funding

There were no funding nor other sources of support to be disclosed. Neither the authors nor any members of their families have received any financial remuneration related to the subject of the article.

Availability of data and materials

The datasets used and analyzed during the current study are available from the corresponding author upon reasonable request.

Declarations

Ethical approval and consent to participate

All animal procedures were performed under the approval and guidance of the Animal Research Committee of Hiroshima University (Approval Number: M19007).

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Received: 29 March 2022 Accepted: 4 October 2022

Published online: 08 October 2022

References

- Maradit Kremers H, Larson DR, Crowson CS, Kremers WK, Washington RE, Steiner CA, et al. Prevalence of total hip and knee replacement in the United States. *J Bone Joint Surg Am*. 2015;97:1386–97. <https://doi.org/10.2106/JBJS.N.01141>.
- Schairer WW, Vail TP, Bozic KJ. What are the rates and causes of hospital readmission after total knee arthroplasty? *Clin Orthop Relat Res*. 2014;472:181–7. <https://doi.org/10.1007/s11999-013-3030-7>.
- Arden NK, Kiran A, Judge A, Biant LC, Javaid MK, Murray DW, et al. What is a good patient reported outcome after total hip replacement? *Osteoarthritis Cartil*. 2011;19:155–62. <https://doi.org/10.1016/j.joca.2010.10.004>.
- Liu X, Chu PK, Ding C. Surface modification of titanium, titanium alloys, and related materials for biomedical applications. *Mater Sci Eng R Rep*. 2004;47:49–121. <https://doi.org/10.1016/j.mser.2004.11.001>.
- Navarro M, Michiardi A, Castaño O, Planell JA. Biomaterials in orthopaedics. *J R Soc Interface*. 2008;5:1137–58. <https://doi.org/10.1098/rsif.2008.0151>.
- Att W, Hori N, Takeuchi M, Ouyang J, Yang Y, Anpo M, et al. Time-dependent degradation of titanium osteoconductivity: an implication of biological aging of implant materials. *Biomaterials*. 2009;30:5352–63. <https://doi.org/10.1016/j.biomaterials.2009.06.040>.
- Puleo DA, Nanci A. Understanding and controlling the bone-implant interface. *Biomaterials*. 1999;20:2311–21. [https://doi.org/10.1016/s0142-9612\(99\)00160-x](https://doi.org/10.1016/s0142-9612(99)00160-x).
- Pilliar RM. Cementless implant fixation—toward improved reliability. *Orthop Clin North Am*. 2005;36:113–9. <https://doi.org/10.1016/j.jocl.2004.08.001>.
- International Diabetes Federation. IDF diabetes atlas. 9th ed. <https://diabetesatlas.org/en/resources/> (date last accessed 1 September 2021)
- Maradit Kremers H, Schleck CD, Lewallen EA, Larson DR, Van Wijnen AJ, Lewallen DG. Diabetes mellitus and hyperglycemia and the risk of aseptic loosening in total joint arthroplasty. *J Arthroplasty*. 2017;32:S251–3. <https://doi.org/10.1016/j.arth.2017.02.056>.
- Maradit Kremers HM, Lewallen LW, Mabry TM, Berry DJ, Berbari EF, Osmon DR. Diabetes mellitus, hyperglycemia, hemoglobin A1C and the risk of prosthetic joint infections in total hip and knee arthroplasty. *J Arthroplasty*. 2015;30:439–43. <https://doi.org/10.1016/j.arth.2014.10.009>.
- Martínez-Huedo MA, Jiménez-García R, Jiménez-Trujillo I, Hernández-Barrera V, Del Rio LB, López-De-Andrés A. Effect of type 2 diabetes on in-hospital postoperative complications and mortality after primary total hip and knee arthroplasty. *J Arthroplasty*. 2017;32:3729–3734.e2. <https://doi.org/10.1016/j.arth.2017.06.038>.
- Morris HF, Ochi S, Winkler S. Implant survival in patients with type 2 diabetes: placement to 36 months. *Ann Periodontol*. 2000;5:157–65. <https://doi.org/10.1902/annals.2000.5.1.157>.
- Olson JW, Shernoff AF, Tarlow JL, Colwell JA, Scheetz JP, Bingham SF. Dental endosseous implant assessments in a type 2 diabetic population: a prospective study. *Int J Oral Maxillofac Implants*. 2000;15:811–8.
- McCracken M, Lemons JE, Rahemtulla F, Prince CW, Feldman D. Bone response to titanium alloy implants placed in diabetic rats. *Int J Oral Maxillofac Implants*. 2000;15:345–54.

16. Fiorellini JP, Nevins ML, Norkin A, Weber HP, Karimbux NY. The effect of insulin therapy on osseointegration in a diabetic rat model. *Clin Oral Implants Res.* 1999;10:362–8. <https://doi.org/10.1111/j.1600-0501.1999.tb00011.x>.
17. Kwon PT, Rahman SS, Kim DM, Kopman JA, Karimbux NY, Fiorellini JP. Maintenance of osseointegration utilizing insulin therapy in a diabetic rat model. *J Periodontol.* 2005;76:621–6. <https://doi.org/10.1902/jop.2005.76.4.621>.
18. Aita H, Hori N, Takeuchi M, Suzuki T, Yamada M, Anpo M, et al. The effect of ultraviolet functionalization of titanium on integration with bone. *Biomaterials.* 2009;30:1015–25. <https://doi.org/10.1016/j.biomaterials.2008.11.004>.
19. Yadav A, Yadav R, Gupta A, Baranwal A, Bhatnagar A, Singh V. Effect of ultraviolet irradiation on the osseointegration of a titanium alloy with bone. *Contemp Clin Dent.* 2017;8:571–8. https://doi.org/10.4103/ccd.ccd_576_17.
20. Minamikawa H, Ikeda T, Att W, Hagiwara Y, Hirota M, Tabuchi M, et al. Photofunctionalization increases the bioactivity and osteoconductivity of the titanium alloy Ti6Al4V. *J Biomed Mater Res A.* 2014;102:3618–30. <https://doi.org/10.1002/jbma.35030>.
21. Yamauchi R, Itabashi T, Wada K, Tanaka T, Kumagai G, Ishibashi Y, et al. Photofunctionalized Ti6Al4V implants enhance early phase osseointegration. *Bone Joint Res.* 2017;6:331–6. <https://doi.org/10.1302/2046-3758.65.BJR-2016-0221.R1>.
22. Hirota M, Tanaka M, Ishijima M, Iwasaki C, Park W, Ogawa T. Effect of photofunctionalization on Ti6Al4V screw stability placed in segmental bone defects in rat femurs. *J Oral Maxillofac Surg.* 2016;74(861):e1-16. <https://doi.org/10.1016/j.joms.2015.11.016>.
23. Sugita Y, Honda Y, Kato I, Kubo K, Maeda H, Ogawa T. Role of Photofunctionalization in mitigating impaired osseointegration associated with type 2 diabetes in rats. *Int J Oral Maxillofac Implants.* 2014;29:1293–300. <https://doi.org/10.11607/jomi.3480>.
24. Masuyama T, Katsuda Y, Shinohara M. A novel model of obesity-related diabetes: introgression of the Lepr^{fa} allele of the Zucker fatty rat into nonobese Spontaneously Diabetic Torii (SDT) rats. *Exp Anim.* 2005;54:13–20. <https://doi.org/10.1538/expanim.54.13>.
25. Murai Y, Sasase T, Tadaki H, Heitaku S, Imagawa N, Yamada T, et al. Analysis of haemodynamics and angiogenic response to ischaemia in the obese type 2 diabetic model Spontaneously Diabetic Torii Lepr^{fa} (SDT fatty) rats. *Clin Exp Pharmacol Physiol.* 2020;47:583–90. <https://doi.org/10.1111/1440-1681.13239>.
26. Katsuda Y, Ohta T, Miyajima K, Kemmochi Y, Sasase T, Tong B, et al. Diabetic complications in obese type 2 diabetic rat models. *Exp Anim.* 2014;63:121–32. <https://doi.org/10.1538/expanim.63.121>.
27. Yamazaki M, Yamada M, Ishizaki K, Sakurai K. Ultraviolet-C irradiation to titanium implants increases peri-implant bone formation without impeding mineralization in a rabbit femur model. *Acta Odontol Scand.* 2015;73:302–11. <https://doi.org/10.3109/00016357.2014.956332>.
28. Oates TW, Dowell S, Robinson M, McMahan CA. Glycemic control and implant stabilization in type 2 diabetes mellitus. *J Dent Res.* 2009;88:367–71. <https://doi.org/10.1177/0022034509334203>.
29. Kilpadi DV, Lemons JE, Liu J, Raikar GN, Weimer JJ, Vohra Y. Cleaning and heat-treatment effects on unalloyed titanium implant surfaces. *Int J Oral Maxillofac Implants.* 2000;15:219–30.
30. Hori N, Att W, Ueno T, Sato N, Yamada M, Saruwatari L, et al. Age-dependent degradation of the protein adsorption capacity of titanium. *J Dent Res.* 2009;88:663–7. <https://doi.org/10.1177/0022034509339567>.
31. Hayashi R, Ueno T, Migita S, Tsutsumi Y, Doi H, Ogawa T, et al. Hydrocarbon deposition attenuates osteoblast activity on titanium. *J Dent Res.* 2014;93:698–703. <https://doi.org/10.1177/0022034514536578>.
32. Wei J, Yoshinari M, Takemoto S, Hattori M, Kawada E, Liu B, et al. Adhesion of mouse fibroblasts on hexamethyldisiloxane surfaces with wide range of wettability. *J Biomed Mater Res B Appl Biomater.* 2007;81:66–75. <https://doi.org/10.1002/jbm.b.30638>.
33. Groth T, Altankov G. Studies on cell-biomaterial interaction: role of tyrosine phosphorylation during fibroblast spreading on surfaces varying in wettability. *Biomaterials.* 1996;17:1227–34. [https://doi.org/10.1016/0142-9612\(96\)84943-x](https://doi.org/10.1016/0142-9612(96)84943-x).
34. Altankov G, Grinnell F, Groth T. Studies on the biocompatibility of materials: fibroblast reorganization of substratum-bound fibronectin on surfaces varying in wettability. *J Biomed Mater Res.* 1996;30:385–91. [https://doi.org/10.1002/\(SICI\)1097-4636\(199603\)30:3<385::AID-JBM13>3.0.CO;2-J](https://doi.org/10.1002/(SICI)1097-4636(199603)30:3<385::AID-JBM13>3.0.CO;2-J).
35. Pyo SW, Park YB, Moon HS, Lee JH, Ogawa T. Photofunctionalization enhances bone-implant contact, dynamics of interfacial osteogenesis, marginal bone seal, and removal torque value of implants: a dog jawbone study. *Implant Dent.* 2013;22:666–75. <https://doi.org/10.1097/ID.0000000000000003>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

